

Service  
Training  
Calibration  
Certification  
Maintenance

Universal Thermal Services  
Allen R. Hildebrand (President)  
E7064 Knopp Road  
Manawa, Wisconsin 54949  
Phone -- Fax 920-596-2983  
uts@wolfnet.net

## Metco 9MC Certificate of Calibration

Form Instructions: F-310-016 Rev. AE 31-Dec-13

Work Instructions: WI-016 Rev. A 22-Feb-10

Certificate Number: 2024-418

Page: 1 of 7

Customer:	Universal Thermal	Console:	Metco 9MC	Powder Feeder:	N/A
Address:	217 Center St	Serial Number:	9MC KM92570231-2	Serial Number:	N/A
City:	Manawa	Device ID Number:	1455		
State:	WI	Booth Number:	N/A	Powder Feeder:	N/A
Zip:	54949	Calibration As Received		Serial Number:	N/A
Name:	Allen Hildebrand	With in 3%:	No		
Phone Number:	920-596-2983	Adjustments:	Yes		

### N.I.S.T. Instruments Used for This Calibration

Test Instrument:	Press. Transducer	Test Instrument:	Multi Meter	Test Instrument:	Amp Clamp Meter	Test Instrument:	Multi Meter
Make:	Fluke	Make:	Fluke	Make:	Fluke	Make:	Fluke
Model:	PV350	Model:	87V	Model:	i1010	Model:	233
Serial Number:	PM-21	Serial Number:	MM-13	Serial Number:	AC-44	Serial Number:	MM-15
Next Calibration Due:	12-Dec-24	Next Calibration Due:	24-May-24	Next Calibration Due:	24-May-24	Next Calibration Due:	7-Dec-24

Test Inst: Low Flow	Mass Flow Meter	Test Instrument: Medium	Mass Flow Meter		
Make:	Alicat 75 scfh	Make:	Alicat 400 scfh		
Model:	PCU50SLPM	Model:	PCU250SLPM		
Serial Number:	282154	Serial Number:	282155		
Next Calibration Due:	7-Sep-24	Next Calibration Due:	7-Sep-24		

### Manufacture Specifications

Flow Tube	Accuracy	Tube Pressure	Tube Scale	Scfh Per Metco	In The % of Deviation Column of Each Sheet After This Sheet Is The Manufacture Specifications, or Better, For The Individual Device
Argon flow meter	± 3%	75 psi	0-300 scfh	300	<b>0-3% Green is Acceptable</b> <b>3.1-5% Orange is Alert</b> <b>5.1% -&gt; Red Is Fail</b>
Hydrogen flow meter	± 3%	50 psi	0-42 scfh	42	
Helium flow meter	± 3%	75 psi	0-225 scfh	225	
Nitrogen flow in Argon Meter	± 3%	75 psi	0-300 scfh	300	

### System Performance w/Parameters as Provided by the Operator Initially

Gas	Pressure	Gas [Ball] Flow	System	Amps	Volts			
Ar	80	100	System	500	37.5			
H2	50	10	Gun Type					
N2	N/A	N/A	9MB					
He	N/A	N/A						

**Notes:**

Primary Flow Meter was Leaking, Found glass flow tube was chipped, Replaced Flow tube, float, and flow meter scale.  
Replaced Hydrogen Flow Meter Scale.


Console Air Purge Safety Test: **Pass**  
 Door Switch Safety Test: **Pass**  
 Argon Supply Test: **Pass**  
 Hydrogen Supply Test: **Pass**  
 Nitrogen Supply Test: **Pass**  
 Helium Supply Test: **Pass**  
 E-Stop Console: **Pass**  
 Leak Test: **Pass**

Calibrated By: Jeremy Bailey  
 Calibrated Date: 15-Feb-24

Calibration Due: 15-Feb-25

Signature:

**All instruments have been calibrated against standards traceable to NIST. This Certification Sheet must not be altered in any way!**



Service  
Training  
Calibration  
Certification  
Maintenance

Universal Thermal Services  
Allen R. Hildebrand (President)  
E7064 Knopp Road  
Manawa, Wisconsin 54949  
Phone – Fax 920-596-2983  
uts@wolfnet.net

Customer: Universal Thermal

## Amperage and Voltage Meter

Form Instructions: F-310-016 Rev. AE 31-Dec-13

Certificate Number : 2024-418

Page: 2 of 7

Console: Metco 9MC Serial Number: 9MC KM92570231-2 Booth : N/A

Testing Instrument: Amp Clamp Meter i1010 Serial Number: AC-44 Testing Instrument: Multi Meter 87V Serial Number: MM-13

### Amp Meter

Device Under Test: Digital Amp Meter Amps Full Scale [FS] 1000 Device ID Number: N/A

Amps Set Point	Amp Meter Display Reading	As Found NIST Meter	As Found Amps Deviation	As Found Amps % Deviation	Tolerances Pass/Fail	As Left NIST Meter	As Left Amps Deviation	As Left Amps % Deviation	Tolerances Pass/Fail
300	300	308	8.0	0.8	Pass	308	8.0	0.8	Pass
400	400	408	8.0	0.8	Pass	408	8.0	0.8	Pass
500	500	508	8.0	0.8	Pass	508	8.0	0.8	Pass
600	600	610	10.0	1.0	Pass	610	10.0	1.0	Pass
700	700	710	10.0	1.0	Pass	710	10.0	1.0	Pass
800	800	810	10.0	1.0	Pass	810	10.0	1.0	Pass
900	900	909	9.0	0.9	Pass	909	9.0	0.9	Pass

### Volt Meter

Device Under Test: Digital Volt Meter Volts Full Scale [FS] 100 Device ID Number:

Volt Meter Set Point	Volt Meter Display Reading	As Found NIST Meter	As Found Volts Deviation	As Found Volts % Deviation	Tolerances Pass/Fail	As Left NIST Meter	As Left Volts Deviation	As Left Volts % Deviation	Tolerances Pass/Fail
26.0	26.0	26.5	0.5	0.5	Pass	26.5	0.5	0.5	Pass
31.0	31.0	31.8	0.8	0.8	Pass	31.8	0.8	0.8	Pass
40.0	40.0	40.5	0.5	0.5	Pass	40.5	0.5	0.5	Pass
50.0	50.0	50.5	0.5	0.5	Pass	50.5	0.5	0.5	Pass

### System Performance w/Parameters after Calibration

Gas	Pressure	Gas [Ball] Flow	Amps
Ar	80	100	500
H2	50	10	Volts
N2	N/A	N/A	37.5
He	N/A	N/A	


Calibrated By: Jeremy Bailey

Calibrated Date: 15-Feb-24

Calibration Due: 15-Feb-25

Signature:

All instruments have been calibrated against standards traceable to NIST. This Certification Sheet must not be altered in any way!



Service  
Training  
Calibration  
Certification  
Maintenance

Universal Thermal Services  
Allen R. Hildebrand (President)  
E7064 Knopp Road  
Manawa, Wisconsin 54949  
Phone – Fax 920-596-2983  
uts@wolfnet.net

Customer Universal Thermal

## Air Gauges

Form Number: F-310-016 Rev. AE 31-Dec-13

Certificate Number: 2024-418

Page: 3 of 7

Console: Metco 9MC Serial Number: 9MC KM92570231-2 Booth: N/A

Testing Instrument: Press. Transducer PV350 Serial Number: PM-21 Testing Instrument: Multi Meter 87V Serial Number: MM-13

### Vibrator Air #1

Device Under Test: Vibrator #1 Gauge Pressure Full Scale [FS] 60 PSI Device ID Number: N/A

Pressure (PSI) Set Point	As Set Gauge Setting	As Found NIST Pressure	As Found PSI Deviation	As Found PSI % of Deviation	Tolerances Pass/Fail	As Left NIST Pressure	As Left PSI Deviation	As Left PSI % Deviation	Tolerances Pass/Fail
10	10	9.7	-0.3	-0.5	Pass	9.7	-0.3	-0.5	Pass
20	20	20.5	0.5	0.8	Pass	20.5	0.5	0.8	Pass
30	30	30.5	0.5	0.8	Pass	30.5	0.5	0.8	Pass
40	40	41.0	1.0	1.7	Pass	41.0	1.0	1.7	Pass
50	50	51.0	1.0	1.7	Pass	51.0	1.0	1.7	Pass
60	60	61.0	1.0	1.7	Pass	61.0	1.0	1.7	Pass

### Vibrator Air #2

Device Under Test: Vibrator #2 Gauge Pressure Full Scale [FS] 60 PSI Device ID Number: N/A

Pressure (PSI) Set Point	As Set Gauge Setting	As Found NIST Pressure	As Found PSI Deviation	As Found PSI % of Deviation	Tolerances Pass/Fail	As Left NIST Pressure	As Left PSI Deviation	As Left PSI % Deviation	Tolerances Pass/Fail
10	10	10.3	0.3	0.5	Pass	10.3	0.3	0.5	Pass
20	20	20.3	0.3	0.5	Pass	20.3	0.3	0.5	Pass
30	30	30.3	0.3	0.5	Pass	30.3	0.3	0.5	Pass
40	40	40.7	0.7	1.2	Pass	40.7	0.7	1.2	Pass
50	50	51.0	1.0	1.7	Pass	51.0	1.0	1.7	Pass
60	60	61.2	1.2	2.0	Pass	61.2	1.2	2.0	Pass

### Cooling Air

Device Under Test: Cooling Air Gauge Pressure Full Scale [FS] 160 PSI Device ID Number: N/A

Pressure (PSI) Set Point	As Set Gauge Setting	As Found NIST Pressure	As Found PSI Deviation	As Found PSI % of Deviation	Tolerances Pass/Fail	As Left NIST Pressure	As Left PSI Deviation	As Left PSI % Deviation	Tolerances Pass/Fail
20	20	20.7	0.7	0.4	Pass	20.7	0.7	0.4	Pass
40	40	41.2	1.2	0.8	Pass	41.2	1.2	0.8	Pass
60	60	61.0	1.0	0.6	Pass	61.0	1.0	0.6	Pass
80	80	81.3	1.3	0.8	Pass	81.3	1.3	0.8	Pass
						0.0			


Calibrated By: Jeremy Bailey

Calibrated Date: 15-Feb-24

Calibration Due: 15-Feb-25

Signature: 

All instruments have been calibrated against standards traceable to NIST. This Certification Sheet must not be altered in any way!



Service  
Training  
Calibration  
Certification  
Maintenance

Universal Thermal Services  
Allen R. Hildebrand (President)  
E7064 Knopp Road  
Manawa, Wisconsin 54949  
Phone – Fax 920-596-2983  
uts@wolfnet.net

Customer: Universal Thermal

## Gas Gauges

Form Number: F-310-016 Rev. AE 31-Dec-13  
Page: 4 of 7

Certificate Number : 2024-418

Console:	Metco 9MC	Serial Number:	9MC KM92570231-2	Booth :	N/A
Testing Instrument:	Press. Transducer PV350	Serial Number:	PM-21	Testing Instrument:	Multi Meter 87V
				Serial Number:	MM-13

### Primary Gas Pressure Gauge

Device Under Test:	Arc Gas Gauge	Pressure Full Scale [FS]	100 PSI	Device ID Number:	N/A
--------------------	---------------	--------------------------	---------	-------------------	-----

Pressure (PSI) Set Point	As Set Gauge Setting	As Found NIST Pressure	As Found PSI Deviation	As Found PSI % of Deviation	Tolerences Pass/Fail	As Left NIST Pressure	As Left PSI Deviation	As Left PSI % Deviation	Tolerences Pass/Fail
40	40	40.8	0.8	0.8	Pass	40.8	0.8	0.8	Pass
50	50	50.5	0.5	0.5	Pass	50.5	0.5	0.5	Pass
60	60	60.5	0.5	0.5	Pass	60.5	0.5	0.5	Pass
70	70	70.2	0.2	0.2	Pass	70.2	0.2	0.2	Pass
80	80	80.6	0.6	0.6	Pass	80.6	0.6	0.6	Pass
90	90	91.0	1.0	1.0	Pass	91.0	1.0	1.0	Pass
100	100	101.7	1.7	1.7	Pass	101.7	1.7	1.7	Pass

### Secondary Gas Pressure Gauge

Device Under Test:	Sec. Gas Gauge	Pressure Full Scale [FS]	100 PSI	Device ID Number:	N/A
--------------------	----------------	--------------------------	---------	-------------------	-----

Pressure (PSI) Set Point	As Set Gauge Setting	As Found NIST Pressure	As Found PSI Deviation	As Found PSI % of Deviation	Tolerences Pass/Fail	As Left NIST Pressure	As Left PSI Deviation	As Left PSI % Deviation	Tolerences Pass/Fail
40	40	40.1	0.1	0.1	Pass	40.1	0.1	0.1	Pass
50	50	50.0	0.0	0.0	Pass	50.0	0.0	0.0	Pass
60	60	60.1	0.1	0.1	Pass	60.1	0.1	0.1	Pass
70	70	70.1	0.1	0.1	Pass	70.1	0.1	0.1	Pass
80	80	80.0	0.0	0.0	Pass	80.0	0.0	0.0	Pass
90	90	90.0	0.0	0.0	Pass	90.0	0.0	0.0	Pass
100	100	99.8	-0.2	-0.2	Pass	99.8	-0.2	-0.2	Pass


Calibrated By: Jeremy Bailey

Calibrated Date: 15-Feb-24

Calibration Due: 15-Feb-25

Signature:

**All instruments have been calibrated against standards traceable to NIST. This Certification Sheet must not be altered in any way!**



Service  
Training  
Calibration  
Certification  
Maintenance

Universal Thermal Services  
Allen R. Hildebrand (President)  
E7064 Knopp Road  
Manawa, Wisconsin 54949  
Phone -- Fax 920-596-2983  
uts@wolfnet.net

Customer: Universal Thermal

## Argon Flow

Tube Scale 0-300 @ 75 psi Ar @ 70°F

Form Number: F-310-016 Rev. AE 31-Dec-13


Certificate Number: 2024-418

Page: 5 of 7

Console:	Metco 9MC	Serial Number:	None
Device ID Number:		Booth:	None
Device Under Test:	Primary Flow Meter	Type of Gas:	Argon
		Scale Rate:	Flow is in SCFH
		Full Flow SCFH [FS]:	300

Testing Instrument: Mass Flow Meter    Alicat 400 scfh    Serial Number: 282155


Flow Tube SCFH Set Point	Flow Tube SCFH Converted	As Found NIST SCFH	As Found Actual SCFH Deviation	As Found SCFH % Deviation	Tolerences Pass/Fail	Adjusted As Left NIST SCFH	As Left SCFH Deviation	As Left SCFH % Deviation	Tolerences Pass/Fail
40	40	52.2	12.2	4.1	Pass	44.8	4.8	1.6	Pass
70	70	81.3	11.3	3.8	Pass	73.4	3.4	1.1	Pass
100	100	109.8	9.8	3.3	Pass	102.4	2.4	0.8	Pass
130	130	139.2	9.2	3.1	Pass	133.4	3.4	1.1	Pass
160	160	167.2	7.2	2.4	Pass	160.8	0.8	0.3	Pass
190	190	193.3	3.3	1.1	Pass	188.0	-2.0	-0.7	Pass
220	220	221.3	1.3	0.4	Pass	217.7	-2.3	-0.8	Pass
250	250	247.2	-2.8	-0.9	Pass	245.6	-4.4	-1.5	Pass



Calibrated By: Jeremy Bailey  
Calibrated Date: 15-Feb-24

Calibration Due: 15-Feb-25    Signature: *[Handwritten Signature]*

All instruments have been calibrated against standards traceable to NIST. This Certification Sheet must not be altered in any way!



Service  
Training  
Calibration  
Certification  
Maintenance

Universal Thermal Services  
Allen R. Hildebrand (President)  
E7064 Knopp Road  
Manawa, Wisconsin 54949  
Phone -- Fax 920-596-2983  
uts@wolfnet.net

Customer: Universal Thermal

## Hydrogen Flow

**Tube Scale 0-42 @ 50 psi H2 @ 70 ° F**

Form Number: F-310-016 Rev. AE 31-Dec-13


Certificate Number: 2024-418

Page: 6 of 7

Console:	Metco 9MC	Serial Number:	9MC KM92570231-2		
Device ID Number:	N/A	Booth:	N/A		
Device Under Test:	Secondary Gas Flow Tube	Type of Gas:	Hydrogen	Scale Rate:	Flow is in SCFH
				Full Flow SCFH [FS]:	42

Testing Instrument:	Mass Flow Meter	Alicat 75 scfh	Serial Number:	282154
---------------------	-----------------	----------------	----------------	--------

Flow Tube SCFH Set Point	Flow Tube SCFH Converted	As Found NIST SCFH	As Found Actual SCFH Deviation	As Found SCFH % Deviation	Tolerances Pass/Fail	Adjusted		As Left SCFH Deviation	As Left SCFH % Deviation	Tolerances Pass/Fail
						As Left NIST SCFH	As Left SCFH			
5	5	3.5	-1.5	-3.6	Pass	5.1	0.1	0.2	Pass	
10	10	8.5	-1.5	-3.6	Pass	10.0	0.0	0.0	Pass	
15	15	13.7	-1.3	-3.1	Pass	14.9	-0.1	-0.2	Pass	
20	20	18.7	-1.3	-3.1	Pass	20.0	0.0	0.0	Pass	
25	25	24.3	-0.7	-1.7	Pass	25.8	0.8	1.9	Pass	
30	30	29.2	-0.8	-1.9	Pass	30.5	0.5	1.2	Pass	
35	35	34.2	-0.8	-1.9	Pass	35.6	0.6	1.4	Pass	
40	40	39.2	-0.8	-1.9	Pass	40.5	0.5	1.2	Pass	




	1	2	3	4	5	6	7	8
Flow SCFH As Found	3.5	8.5	13.7	18.7	24.3	29.2	34.2	39.2
Flow SCFH As Left	5.1	10.0	14.9	20.0	25.8	30.5	35.6	40.5
SCFH Set Points	5	10	15	20	25	30	35	40

Calibrated By: Jeremy Bailey  
Calibrated Date: 15-Feb-24

Calibration Due: 15-Feb-25

Signature: *[Handwritten Signature]*

**All instruments have been calibrated against standards traceable to NIST. This Certification Sheet must not be altered in any way!**



Service  
Training  
Calibration  
Certification  
Maintenance

Universal Thermal Services  
Allen R. Hildebrand (President)  
E7064 Knopp Road  
Manawa, Wisconsin 54949  
Phone -- Fax 920-596-2983  
uts@wolfnet.net

Customer: Universal Thermal

## Helium Flow

Tube Scale 0-225 @ 75 psi He @ 70°F

Form Number: F-310-016 Rev. AE 31-Dec-13


Certificate Number: 2024-418

Page: 7 of 7

Console:	Metco 9MC	Serial Number:	9MC KM92570231-2
Device ID Number:	N/A	Booth:	N/A
Device Under Test:	Secondary Gas Flow Tube	Type of Gas:	Helium
		Scale Rate:	Flow is in SCFH
		Full Flow SCFH [FS]:	225

Testing Instrument:	Mass Flow Meter	Alicat 400 scfh	Serial Number:	282155
---------------------	-----------------	-----------------	----------------	--------

Flow Tube SCFH Set Point	Flow Tube SCFH Converted	As Found NIST SCFH	As Found Actual SCFH Deviation	As Found SCFH % Deviation	Tolerances Pass/Fail	As Left NIST SCFH	As Left SCFH Deviation	As Left SCFH % Deviation	Tolerances Pass/Fail
40	40	42.6	2.6	1.2	Pass	42.6	2.6	1.2	Pass
60	60	63.1	3.1	1.4	Pass	63.1	3.1	1.4	Pass
80	80	82.7	2.7	1.2	Pass	82.7	2.7	1.2	Pass
100	100	102.8	2.8	1.2	Pass	102.8	2.8	1.2	Pass
120	120	122.7	2.7	1.2	Pass	122.7	2.7	1.2	Pass
140	140	143.8	3.8	1.7	Pass	143.8	3.8	1.7	Pass
160	160	162.8	2.8	1.2	Pass	162.8	2.8	1.2	Pass
180	180	182.8	2.8	1.2	Pass	182.8	2.8	1.2	Pass
200	200	203.8	3.8	1.7	Pass	203.8	3.8	1.7	Pass



	1	2	3	4	5	6	7	8	9
Flow SCFH As Found	42.6	63.1	82.7	102.8	122.7	143.8	162.8	182.8	203.8
Flow SCFH As Left	42.6	63.1	82.7	102.8	122.7	143.8	162.8	182.8	203.8
SCFH Set Points	40	60	80	100	120	140	160	180	200

Calibrated By: Jeremy Bailey  
Calibrated Date: 15-Feb-24

Calibration Due: 15-Feb-25  
Signature: *[Handwritten Signature]*

All instruments have been calibrated against standards traceable to NIST. This Certification Sheet must not be altered in any way!